

# ABOUT YOUR HEALTH SERVICES

Whole body NLS diagnosis & therapy

H/no. A/59 AwutuBreku, P O Box 3463 KaneshieAccra

Tel:0241-262339/0241-169761/0278-547926

THE CHAIRMAN

CIT NELIAN, BOLSHAYA SEMENOVSKAYA STREET 40

BUSINESS CENTRE, AGAT, BUILDING 1A

OFFICE 204

107023

MOSCOW RUSSIA

10<sup>th</sup> May 2015

Dear Sir,

## USER REPORT

Name of Device: DIANEL – 22S – iON

Serial No: 0100097

Date: February 2014

Organisation: ABOUT YOUR HEALTH SERVICES

Focus of Organisation: Patients Health Care

I am a registered Medical practitioner with the Medical and Dental Council of Ghana,  
Registration Number: MDC/RN/0645 and the business is registered with the Accountant  
Generals Department of Ghana, Registration Number: BN-44-533D.

# ABOUT YOUR HEALTH SERVICES

Whole body NLS diagnosis & therapy

H/no. A/59 AwutuBreku, P O Box 3463 KaneshieAccrra

Tel:0241-262339/0241-169761/0278-547926

Using the operation of the device from February 2014 to May 2015 the total number of patients seen is **626**.

**DISEASES DIAGNOSED:** The device has diagnosed mainly cardiovascular diseases and stroke (cardiovascular accidents), gastrointestinal infections and inflammation, gastric ulcer, joint and skeletal problems, insomnia (poor sleep), mental or psychological problems, liver disease especially hepatitis B, headache and migraine, fibroid and infertility in females.

**ACCURACY OF DEVICE:** Approximately 60% matched clinical diagnosis and 40% were newly diagnosed by the device. On a few diagnostic sessions (about 10-15% of the time) the device has wrongly diagnosed the patient's problem. The Dianel device is a highly efficient and accurate diagnostic tool and the use of it to find medicinal preparations have been quite good. Occasionally I have supplemented the herbal preparations with pharmaceutical products

**THE DEVICE:** The Dianel device is very user friendly and the prompts respond to commands very well. When I use it in my clinic I take about 20-30 minutes with each patient on average, slightly longer with older patients because of the chronic nature of their complaints.

On a very few occasions I have had to restart the computer because of hard disk issues and I have added an external hard disk to the laptop to improve operation and storage of

# ABOUT YOUR HEALTH SERVICES

Whole body NLS diagnosis & therapy

H/no. A/59 AwutuBreku, P O Box 3463 KaneshieAccra

Tel:0241-262339/0241-169761/0278-547926

information. There are times when the device will cease scanning for about 3 -4 seconds before it continues. The only other problems I have had with the device is the need for frequent license renewal, otherwise it is a reliable device.

Yours Faithfully,



Dr. Samuel K Acquah

BDS, MSc, MScDPH

Proprietor

